

Certificate of Analysis

Product name:

Urodiag[®] Multiplex PCR Kit

Cat #: UR50P

Lot #: 00001

Storage Condition: -20°C

Expiration Date: May 2021

Quality Control: The above lot has been tested as described in the Multiplex PCR protocol of the Urodiag handbook available on www.oncodiag.fr.

Results: It is certified that the Ct values meet specifications and approved by the Quality Assurance Manager.

Mutation Assay					
Control C+ mut 1 (Ct values)			Control C+ mut 2 (Ct values)		
GLOBIN	S249C	Y375C	GLOBIN	R248C	G372C
30.20	31.82	30.51	30.11	30.83	30.24

Methylation Assay			
Control C+ meth (Ct values)			
ALBUMIN	SEPTIN9	HS3ST2	SLIT2
27.38	26.95	27.17	27.57



Hanane Boucheniata
Quality Assurance Manager

May 20, 2020

Date